## PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

## Procurement of GOODS: Supply and Installation of ATOMIC ABSORPTION SPECTROMETER

Government of the Republic of the Philippines
PHILIPPINE SCIENCE HIGH SCHOOL
MAIN CAMPUS
Agham Road, Diliman, Quezon City

Sixth Edition July 2020

#### **Preface**

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Unit. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents shall clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Care should be taken to check the relevance of the provisions of the PBDs against the requirements of the specific Goods to be procured. If duplication of a subject is inevitable in other sections of the document prepared by the Procuring Entity, care must be exercised to avoid contradictions between clauses dealing with the same matter.

Moreover, each section is prepared with notes intended only as information for the Procuring Entity or the person drafting the Bidding Documents. They shall not be included in the final documents. The following general directions should be observed when using the documents:

- a. All the documents listed in the Table of Contents are normally required for the procurement of Goods. However, they should be adapted as necessary to the circumstances of the particular Procurement Project.
- b. Specific details, such as the "name of the Procuring Entity" and "address for bid submission," should be furnished in the Instructions to Bidders, Bid Data Sheet, and Special Conditions of Contract. The final documents should contain neither blank spaces nor options.
- c. This Preface and the footnotes or notes in italics included in the Invitation to Bid, Bid Data Sheet, General Conditions of Contract, Special Conditions of Contract, Schedule of Requirements, and Specifications are not part of the text of the final document, although they contain instructions that the Procuring Entity should strictly follow.

- d. The cover should be modified as required to identify the Bidding Documents as to the Procurement Project, Project Identification Number, and Procuring Entity, in addition to the date of issue.
- e. Modifications for specific Procurement Project details should be provided in the Special Conditions of Contract as amendments to the Conditions of Contract. For easy completion, whenever reference has to be made to specific clauses in the Bid Data Sheet or Special Conditions of Contract, these terms shall be printed in bold typeface on Sections I (Instructions to Bidders) and III (General Conditions of Contract), respectively.
- f. For guidelines on the use of Bidding Forms and the procurement of Foreign-Assisted Projects, these will be covered by a separate issuance of the Government Procurement Policy Board.

### **Table of Contents**

Gloss	ary of Acronyms, Terms, and Abbreviations	4
Sectio	on I. Invitation to Bid	7
Sectio	on II. Instructions to Bidders	10
1.	Scope of Bid	11
2.	Funding Information	11
3.	Bidding Requirements	11
4.	Corrupt, Fraudulent, Collusive, and Coercive Practices	11
5.	Eligible Bidders	12
6.	Origin of Goods	12
7.	Subcontracts	12
8.	Pre-Bid Conference	13
9.	Clarification and Amendment of Bidding Documents	13
10.	Documents comprising the Bid: Eligibility and Technical Components	13
11.	Documents comprising the Bid: Financial Component	14
12.	Bid Prices	14
13.	Bid and Payment Currencies	15
14.	Bid Security	15
15.	Sealing and Marking of Bids	15
16.	Deadline for Submission of Bids	15
17.	Opening and Preliminary Examination of Bids	15
18.	Domestic Preference	16
19.	Detailed Evaluation and Comparison of Bids	16
20.	Post-Qualification	16
21.	Signing of the Contract	17
Sectio	on III. Bid Data Sheet	18
Sectio	on IV. General Conditions of Contract	20
1.	Scope of Contract	21
2.	Advance Payment and Terms of Payment	21
3.	Performance Security	21
4.	Inspection and Tests	21
5.	Warranty	22
6.	Liability of the Supplier	
Sectio	on V. Special Conditions of Contract	23
	on VI. Schedule of Requirements	
	n VII. Technical Specifications	
	on VIII. Checklist of Technical and Financial Documents	
	ar varre Checimide di Accimical anu Phiancial Ducumento	••••••

## Glossary of Acronyms, Terms, and Abbreviations

**ABC** – Approved Budget for the Contract.

**BAC** – Bids and Awards Committee.

**Bid** – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

**Bidder** – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

**Bidding Documents** – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

**BIR** – Bureau of Internal Revenue.

**BSP** – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

**CDA** - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

**CIF** – Cost Insurance and Freight.

**CIP** – Carriage and Insurance Paid.

**CPI** – Consumer Price Index.

**DDP** – Refers to the quoted price of the Goods, which means "delivered duty paid."

**DTI** – Department of Trade and Industry.

 $\mathbf{EXW} - \mathbf{Ex}$  works.

**FCA** – "Free Carrier" shipping point.

**FOB** – "Free on Board" shipping point.

**Foreign-funded Procurement or Foreign-Assisted Project**— Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

**Framework Agreement** – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as "Call-Offs," are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

**GFI** – Government Financial Institution.

**GOCC** – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term "related" or "analogous services" shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

**GOP** – Government of the Philippines.

**GPPB** – Government Procurement Policy Board.

**INCOTERMS** – International Commercial Terms.

**Infrastructure Projects** – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national

buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs – Local Government Units.

NFCC - Net Financial Contracting Capacity.

**NGA** – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

**Procurement Project** – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

**PSA** – Philippine Statistics Authority.

**SEC** – Securities and Exchange Commission.

**SLCC** – Single Largest Completed Contract.

**Supplier** – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

**UN** – United Nations.

### Section I. Invitation to Bid

#### **Notes on the Invitation to Bid**

The Invitation to Bid (IB) provides information that enables potential Bidders to decide whether to participate in the procurement at hand. The IB shall be posted in accordance with Section 21.2 of the 2016 revised IRR of RA No. 9184.

Apart from the essential items listed in the Bidding Documents, the IB should also indicate the following:

- a. The date of availability of the Bidding Documents, which shall be from the time the IB is first advertised/posted until the deadline for the submission and receipt of bids;
- b. The place where the Bidding Documents may be acquired or the website where it may be downloaded;
- c. The deadline for the submission and receipt of bids; and
- d. Any important bid evaluation criteria (*e.g.*, the application of a margin of preference in bid evaluation).

The IB should be incorporated in the Bidding Documents. The information contained in the IB must conform to the Bidding Documents and in particular to the relevant information in the Bid Data Sheet.



# Republic of the Philippines Department of Science and Technology PHILIPPINE SCIENCE HIGH SCHOOL MAIN CAMPUS Agham Road, Diliman, Quezon City





## INVITATION TO BID FOR Supply and Installation of Atomic Absorption Spectrometer

- 1. The *PHILIPPINE SCIENCE HIGH SCHOOL MAIN CAMPUS*, through the *GAA 2020* intends to apply the sum of *Two Million Five Hundred Thousand Pesos (Php 2,500,000.00)* being the ABC to payments under the contract for *Supply and Installation of Atomic Absorption Spectrometer*. Bids received in excess of the ABC shall be automatically rejected at bid opening.
- 2. The *PHILIPPINE SCIENCE HIGH SCHOOL MAIN CAMPUS* now invites bids for the above Procurement Project. Delivery of the Goods is required by *one hundred twenty (120) calendar days*. Bidders should have completed, within *five (5) years* from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).
- 3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary "pass/fail" criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.
- 4. Bidding is open to all interested bidders, whether local or foreign, subject to the conditions for eligibility provided in the 2016 revised IRR of RA No. 9184.
- 5. Prospective Bidders may obtain further information from *PHILIPPINE SCIENCE HIGH SCHOOL MAIN CAMPUS* and inspect the Bidding Documents at the address given below during 9:00 a.m. to 3:00 p.m.
- 6. A complete set of Bidding Documents may be acquired by interested Bidders on *October 8, 2020 until before the opening of bids* from the given address and website(s) below and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the amount of *Php 5,000.00*. The Procuring Entity shall allow the bidder to present its proof of payment for the fees in person, by facsimile, or through electronic means.
- 7. The *PHILIPPINE SCIENCE HIGH SCHOOL MAIN CAMPUS* will hold a Pre-Bid Conference<sup>1</sup> on *October 16, 2020 at 3:00 p.m.* through video conferencing or webcasting *via Goggle Meet* (<u>meet.google.com/uhm-yoso-yui)</u>, which shall be open to prospective bidders.

<sup>&</sup>lt;sup>1</sup> May be deleted in case the ABC is less than One Million Pesos (PhP1,000,000) where the Procuring Entity may not hold a Pre-Bid Conference.

- 8. Bids must be duly received by the BAC Secretariat through manual submission at the office address indicated below on or before *October 30, 2020 at 2:45 p.m.* Late bids shall not be accepted.
- 9. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB** Clause 14.
- 10. Bid opening shall be on *October 30, 2020 at 3:00 p.m.* via *Google Meet* (<u>meet.google.com/xos-qxzi-yuj)</u> Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
- 11. The *PHILIPPINE SCIENCE HIGH SCHOOL MAIN CAMPUS* reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.
- 12. For further information, please refer to:

Mr. Azlek Owen B. Tan
Head, BAC Secretariat
Philippine Science High School – Main Campus
Agham Road, Diliman, Quezon City
aobtan@pshs.edu.ph
Telefax No.: 8924-0617 / 7500-1448
www.mc.pshs.edu.ph

13. You may visit the following websites:

For downloading of Bidding Documents: PSHS Main Campus Website http://mc.pshs.edu.ph/goods-and-services/ and/or Philgeps Website

October 7, 2020

(Sgd) Ms. Jennifer C. Bermudez Chairperson, BAC for Goods and Services

## Section II. Instructions to Bidders

#### **Notes on the Instructions to Bidders**

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

#### 1. Scope of Bid

The Procuring Entity, *PHILIPPINE SCIENCE HIGH SCHOOL – MAIN CAMPUS* wishes to receive Bids for the *Supply and Installation of Atomic Absorption Spectrometer*, with identification number 20-10-118.

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

The Procurement Project (referred to herein as "Project") is composed of **One** (1) unit of Atomic Absorption Spectrometer, the details of which are described in Section VII (Technical Specifications).

#### 2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for *GAA 2020* in the amount of *Php 2,500,000.00*.
- 2.2. The source of funding is:
  - a. NGA, the General Appropriations Act or Special Appropriations.

#### 3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

#### 4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

#### 5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
  - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
  - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
  - iii. When the Goods sought to be procured are not available from local suppliers; or
  - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
  - a. For the procurement of Non-expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

#### 6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

#### 7. Subcontracts

7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that:

a. Subcontracting is not allowed.

- 7.2. [If Procuring Entity has determined that subcontracting is allowed during the bidding, state:] The Bidder must submit together with its Bid the documentary requirements of the subcontractor(s) complying with the eligibility criteria stated in **ITB** Clause 5 in accordance with Section 23.4 of the 2016 revised IRR of RA No. 9184 pursuant to Section 23.1 thereof.
- 7.3. [If subcontracting is allowed during the contract implementation stage, state:] The Supplier may identify its subcontractor during the contract implementation stage. Subcontractors identified during the bidding may be changed during the implementation of this Contract. Subcontractors must submit the documentary requirements under Section 23.1 of the 2016 revised IRR of RA No. 9184 and comply with the eligibility criteria specified in ITB Clause 5 to the implementing or end-user unit.
- 7.4. Subcontracting of any portion of the Project does not relieve the Supplier of any liability or obligation under the Contract. The Supplier will be responsible for the acts, defaults, and negligence of any subcontractor, its agents, servants, or workmen as fully as if these were the Supplier's own acts, defaults, or negligence, or those of its agents, servants, or workmen.

#### 8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and **through** *Google Meet* (<u>meet.google.com/uhm-yoso-yui)</u> as indicated in paragraph 6 of the **IB.** 

#### 9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

#### 10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII** (Checklist of Technical and Financial **Documents**).
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *five* (5) *years* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated

through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

#### 11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII** (Checklist of Technical and Financial Documents).
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

#### 12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
  - a. For Goods offered from within the Procuring Entity's country:
    - i. The price of the Goods quoted EXW (ex-works, ex-factory, exwarehouse, ex-showroom, or off-the-shelf, as applicable);
    - ii. The cost of all customs duties and sales and other taxes already paid or payable;
    - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
    - iv. The price of other (incidental) services, if any, listed in e.
  - b. For Goods offered from abroad:
    - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
    - ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications).**

#### 13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
  - a. Philippine Pesos.

#### 14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration<sup>2</sup> or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security shall be valid until *February 27, 2021*. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

#### 15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

#### 16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the **IB**.

#### 17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing,

<sup>&</sup>lt;sup>2</sup> In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

#### 18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

#### 19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "passed," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII** (**Technical Specifications**), although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as follows:
  - Option 1 One Project having several items that shall be awarded as one contract.
- 19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

#### 20. Post-Qualification

20.2. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid.

#### 21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

### Section III. Bid Data Sheet

#### **Notes on the Bid Data Sheet**

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

## **Bid Data Sheet**

ITB		
Clause		
5.3	For this purpose, contracts similar to the Project shall be:	
	a. Supply and Installation of Atomic Absorption Spectrometer.	
	b. completed within <i>five</i> (5) <i>years</i> prior to the deadline for the submission and receipt of bids.	
7.1	"Subcontracting is not allowed."	
12	The price of the Goods shall be quoted DDP <i>Philippine Science High School – Main Campus</i> or the applicable International Commercial Terms (INCOTERMS) for this Project.	
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:	
	a. The amount of not less than <i>Php 50,000.00</i> if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or	
	b. The amount of not less than <i>Php 125,000.00</i> if bid security is in Surety Bond.	
19.3	Not Applicable.	
20.2	"No additional requirement."	
21.2	"No additional requirement."	

## Section IV. General Conditions of Contract

#### **Notes on the General Conditions of Contract**

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

#### 1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC).** 

#### 2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

#### 3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

#### 4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, Section IV (Technical Specifications) shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

#### 5. Warranty

- 6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.
- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

#### 6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

## Section V. Special Conditions of Contract

#### **Notes on the Special Conditions of Contract**

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

**Special Conditions of Contract** 

CCC	Special Conditions of Contract		
GCC Clause			
1	[List here any additional requirements for the completion of this Contract. The following requirements and the corresponding provisions may be deleted amended, or retained depending on its applicability to this Contract:]		
	Delivery and Documents –		
	For purposes of the Contract, "EXW," "FOB," "FCA," "CIF," "CIP," "DDP" and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:		
	[For Goods supplied from abroad, state:] "The delivery terms applicable to the Contract are DDP delivered [indicate place of destination]. In accordance with INCOTERMS."		
	[For Goods supplied from within the Philippines, state:] "The delivery term applicable to this Contract are delivered [indicate place of destination]. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination."		
	Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).		
	For purposes of this Clause the Procuring Entity's Representative at the Project Site is [indicate name(s)].		
	Incidental Services –		
	The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: Select appropriate requirements and delete the rest.		
	<ul><li>a. performance or supervision of on-site assembly and/or start-up of the supplied Goods;</li><li>b. furnishing of tools required for assembly and/or maintenance of the supplied Goods;</li></ul>		
	<ul> <li>c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;</li> <li>d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty</li> </ul>		
	obligations under this Contract; and		

- e. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
- f. [Specify additional incidental service requirements, as needed.]

The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

#### Spare Parts -

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

*Select appropriate requirements and delete the rest.* 

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
  - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
  - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI** (**Schedule of Requirements**) and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of [indicate here the time period specified. If not used indicate a time period of three times the warranty period].

Spare parts or components shall be supplied as promptly as possible, but in any case, within [insert appropriate time period] months of placing the order.

#### Packaging -

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

Name of the Procuring Entity
Name of the Supplier
Contract Description
Final Destination
Gross weight
Any special lifting instructions
Any special handling instructions
Any relevant HAZCHEM classifications

A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.

#### Transportation -

Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.

Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.

Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.

The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.

#### **Intellectual Property Rights –**

The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.

2.2 The terms of payment shall be as follows:

4

The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations.

The supplier may submit a progress billing and corresponding request for progress payment for the delivered goods and/or services accomplished.

The procuring entity shall deduct the following from the certified gross amounts to be paid to the supplier as progress payment:

- All progress payments shall first be charged against the advance payment until the latter has been fully exhausted, unless otherwise approved by the President.
- Retention money in an amount equivalent to at least ten percent (10%) of every progress payment.

The inspections and tests that will be conducted are: The Inspectorate Team with the TWG and/or End User unit shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual.

## Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Quantity	Total	Delivered, Weeks/Months
1	Atomic Absorption Spectrometer	1	1	120 CD

## Section VII. Technical Specifications

#### **Notes for Preparing the Technical Specifications**

A set of precise and clear specifications is a prerequisite for Bidders to respond realistically and competitively to the requirements of the Procuring Entity without qualifying their Bids. In the context of Competitive Bidding, the specifications (*e.g.* production/delivery schedule, manpower requirements, and after-sales service/parts, descriptions of the lots or items) must be prepared to permit the widest possible competition and, at the same time, present a clear statement of the required standards of workmanship, materials, and performance of the goods and services to be procured. Only if this is done will the objectives of transparency, equity, efficiency, fairness, and economy in procurement be realized, responsiveness of bids be ensured, and the subsequent task of bid evaluation and post-qualification facilitated. The specifications should require that all items, materials and accessories to be included or incorporated in the goods be new, unused, and of the most recent or current models, and that they include or incorporate all recent improvements in design and materials unless otherwise provided in the Contract.

Samples of specifications from previous similar procurements are useful in this respect. The use of metric units is encouraged. Depending on the complexity of the goods and the repetitiveness of the type of procurement, it may be advantageous to standardize the General Technical Specifications and incorporate them in a separate subsection. The General Technical Specifications should cover all classes of workmanship, materials, and equipment commonly involved in manufacturing similar goods. Deletions or addenda should then adapt the General Technical Specifications to the particular procurement.

Care must be taken in drafting specifications to ensure that they are not restrictive. In the specification of standards for equipment, materials, and workmanship, recognized Philippine and international standards should be used as much as possible. Where other particular standards are used, whether national standards or other standards, the specifications should state that equipment, materials, and workmanship that meet other authoritative standards, and which ensure at least a substantially equal quality than the standards mentioned, will also be acceptable. The following clause may be inserted in the Special Conditions of Contract or the Technical Specifications.

#### Sample Clause: Equivalency of Standards and Codes

Wherever reference is made in the Technical Specifications to specific standards and codes to be met by the goods and materials to be furnished or tested, the provisions of the latest edition or revision of the relevant standards and codes shall apply, unless otherwise expressly stated in the Contract. Where such standards and codes are national or relate to a particular country or region, other authoritative standards that ensure substantial equivalence to the standards and codes specified will be acceptable.

Reference to brand name and catalogue number should be avoided as far as possible; where unavoidable they should always be followed by the words "or at least equivalent." References to brand names cannot be used when the funding source is the GOP.

Where appropriate, drawings, including site plans as required, may be furnished by the Procuring Entity with the Bidding Documents. Similarly, the Supplier may be requested to provide drawings or samples either with its Bid or for prior review by the Procuring Entity during contract execution.

Bidders are also required, as part of the technical specifications, to complete their statement of compliance demonstrating how the items comply with the specification.

## **Technical Specifications**

Item	Specification	Statement of Compliance
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]
1	Procurement of One (1) unit Atomic Absorption Spectrometer	
	Specifications: Optics	
	Wavelength range:185 to 900 nm	
	Monochromator: Aberration- corrected Czerny-Turner mounting	
	Number of grating grooves: 1800 lines / mm	
	Focal length: 300 mm	

D 1 111 01 02 04 07 10	
Bandwidth: 0.1, 0.2, 0.4, 0.7, 1.0,	
2.0 nm (6-step automatically	
switching)	
_	
Detector: Photomultiplier tube	
1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	
Configuration: Optical double	
beam (with Beam Manager)	
beam (with beam Manager)	
De drawaya da a ma ati an mathad.	
Background correction method:	
BGC-SR (high-speed self-	
reversal method) (185.0 to 900.0	
nm)	
BGC-D2(D2lamp method)	
(185.0 to 430.0 nm)	
(103.0 to 430.0 mm)	
Number of HC lamps 9 lamp	
Number of HC lamps:8-lamp	
turret, 2 lamps simultaneously lit,	
1 for measurement, 1 warming	
up for next measurement 2	
sockets for High- performance	
boosted lamp, litting one	
boosted ramp, fitting one	
Lamp mode: EMISSION, NON-	
BGC, BGC-D2, BGC-SR	
Flame Type: Air-cooled premix	
type	
Burner Head: Titanium 10 cm	
slot (5 cm titanium slot for N2O-	
C2H2 flame available as an	
option)	
N. 1. 1' DEED 'C' D. I	
Nebulizer: PTFE orifice, Pt-Ir	
capillary, ceramic impact bead	
(capable of handling hydrofluoric	
acid)	
,	
Chamber: Engineering Plastics	
Positioning: Lateral/vertical	
manual adjustment	
A 1 12 / 2020	
Angle adjustment:0 to 90°	
Flame type:Air-C2H2flame,	
N2O-C2H2flame	

Data processing	
Parameter Setting: Wizard method	
Measurement mode: Flame continuous method, flame micro Sampling method, furnace method, flame emission method	
Repeat analysis: Up to 20 repetitions. Mean value, standard Deviation (SD) and coefficient of variation (RSD) display	
Automatic exclusion of deviant values by setting SD and %RSD	
Baseline correction: Automatic correction of baseline drift by offset correction in peak height / peak area modes QA/QC: Select whether to continue or discontinue measurements based on results of evaluation on coefficient of correlation, %RSD, ICV.ICB, CCV.CCB, PB, LCS, SPK, PDS, and DUP	
Power Requirements:220 AV, 60 Hz Ambient	
Temperature/humidity: 10 to 35°C, 20 to 80% (less than 70% when Temperature is higher than 30 °C)	
A. With Automatic fuel gas leak detector (when the power is turned on and when the flame is extinguished)	
B. With Automatic Air-NO2 switching as C2H2 flow rate increases.	
C. Flashback detection	

D. Flame monitor (Automatic	
Gas Shut-off)	
E. Sudden power interruption detection and re-ignition safety features	
F. Drain tank level monitor	
G. Mechanism to prevent misuse of the burner.	
H. Flame-retardant materials	
WITH ACCESSORIES INCLUDED:	
1 Unit Hydride Vapor Generator	
Features:	
<ul> <li>Environmental standards prescribe the hydride generation method as one method of Arsenic, Selenium and Antimony analysis.</li> <li>HVG-1 permits the rapid and accurate quantitation of elements such as As, Se, Sb, Sn, Te, Bi and Hg at several-pbb levels.</li> </ul>	
Specifications:	
<ul> <li>Measurement method: continuous flow</li> <li>Sample consumption: 0 to 7 mL/min, variable</li> <li>Reagent consumption: 0 to 2.5 mL/min, variable</li> <li>Carrier Gas: Ar pressure: 0.32 MPa, consumption: 70 mL/min</li> <li>Dimension: 340 W x 220 D</li> </ul>	
ADDITIONAL ACCESSORIES:	
Initial Supply of Argon Gas	
Argon Regulator	
■ Dimension: 340 W x 220 D  ADDITIONAL ACCESSORIES:  Initial Supply of Argon Gas	

1 Set Complete Pe System, Intel core-i5 compatible unit with 4GB RAM and 1 Terabyte hard drive. Multimedia ready with CD-ROM drive and 21.5" LCD color monitor, keyboard and mouse, Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Cadmium, (Cd) 1 pc Camium, (Cd) 1 pc Zinc, (Zn) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, (Admium, Chromium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor  1 Set Hood and Ducting works			
core-i5 compatible unit with 4GB RAM and 1 Terabyte hard drive. Multimedia ready with CD-ROM drive and 21.5" LCD color monitor, keyboard and mouse, Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 Set Complete Pc System, Intel	
RAM and 1 Terabyte hard drive. Multimedia ready with CD-ROM drive and 21.5" LCD color monitor, keyboard and mouse, Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 M\Omega.cm @25^C Product Water TOC (with UV): <5ppb Particulate count (>0.22um): <1 particulate/ml Bacteria Count: <0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Cadmium, (Cd) 1 pc Cadmium, (Cd) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		- · · · · · · · · · · · · · · · · · · ·	
Multimedia ready with CD-ROM drive and 21.5" LCD color monitor, keyboard and mouse, Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM 1.  Ultrapure Water Quality (Type 1)  Product Water Resistivity :18.2 MΩ.cm @25°C  Product Water TOC (with UV): <5ppb  Particulate count (>0.22um): < 1 particulate/ ml  Bacteria Count: <0.1 cfu/ml Flow Rate: L/min (Max)  Hallow Cathode Lamp  1 pc Lead (Pb)  1 pc Iron (Fe)  1 pc Magnesium, (Mg)  1 pc Magnesium, (Cd)  1 pc Cadmium, (Cd)  1 pc Chromium, (Cr)  1 pc Arsenic, (As)  1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		•	
drive and 21.5" LCD color monitor, keyboard and mouse, Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
monitor, keyboard and mouse, Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1)  Product Water Resistivity :18.2  MΩ cm @25°C  Product Water TOC (with UV): < 5ppb  Particulate count (>0.22um): < 1 particulate/ ml  Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max)  Hallow Cathode Lamp  1 pc Lead (Pb)  1 pc Iron (Fe)  1 pc Manganese, (Mn)  1 pc Cadmium, (Cd)  1 pc Cadmium, (Cd)  1 pc Sodium, (Na)  1 pc Zinc, (Zn)  1 pc Arsenic, (As)  1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		drive and 21.5" LCD color	
Windows 10 OS.  1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		monitor, keyboard and mouse.	
1 Unit Deskjet Ink Printer, continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		The state of the s	
continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Caromium, (Cr) 1 pc Sodium, (Na) 1 pc Sodium, (Na) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		Willdows 10 OS.	
continuous  1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Caromium, (Cr) 1 pc Sodium, (Na) 1 pc Sodium, (Na) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		444.5.4.4.5.	
1 Unit WATER PURIFICATION SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		continuous	
SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
SYSTEM I.  Ultrapure Water Quality (Type 1) Product Water Resistivity :18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 Unit WATER PURIFICATION	
Ultrapure Water Quality (Type 1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		SISILMI.	
1) Product Water Resistivity:18.2 MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1114 W-4 O 114 (T	
Product Water Resistivity:18.2  MΩ.cm @25°C  Product Water TOC (with UV): < 5ppb  Particulate count (>0.22um): < 1 particulate/ ml  Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max)  Hallow Cathode Lamp  1 pc Lead (Pb)  1 pc Iron (Fe)  1 pc Magnesium, (Mg)  1 pc Manganese, (Mn)  1 pc Cadmium, (Cd)  1 pc Chromium, (Cr)  1 pc Sodium, (Na)  1 pc Zinc, (Zn)  1 pc Arsenic, (As)  1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1)	
MΩ.cm @25°C Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		Product Water Resistivity:18.2	
Product Water TOC (with UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Cadmium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		•	
UV): < 5ppb Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		$\overline{\mathcal{C}}$	
Particulate count (>0.22um): < 1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor	[	· ·	
1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		,	
1 particulate/ ml Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor	•	Particulate count (>0.22um): <	
Bacteria Count: < 0.1 cfu/ml Flow Rate: L/min (Max)  Hallow Cathode Lamp  1 pc Lead (Pb)  1 pc Iron (Fe)  1 pc Magnesium, (Mg)  1 pc Manganese, (Mn)  1 pc Cadmium, (Cd)  1 pc Chromium, (Cr)  1 pc Sodium, (Na)  1 pc Zinc, (Zn)  1 pc Arsenic, (As)  1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		` ,	
Flow Rate: L/min (Max) Hallow Cathode Lamp 1 pc Lead (Pb) 1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1	
Hallow Cathode Lamp  1 pc Lead (Pb)  1 pc Iron (Fe)  1 pc Magnesium, (Mg)  1 pc Manganese, (Mn)  1 pc Cadmium, (Cd)  1 pc Chromium, (Cr)  1 pc Sodium, (Na)  1 pc Zinc, (Zn)  1 pc Arsenic, (As)  1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
<ul> <li>1 pc Lead (Pb)</li> <li>1 pc Iron (Fe)</li> <li>1 pc Magnesium, (Mg)</li> <li>1 pc Manganese, (Mn)</li> <li>1 pc Cadmium, (Cd)</li> <li>1 pc Chromium, (Cr)</li> <li>1 pc Sodium, (Na)</li> <li>1 pc Zinc, (Zn)</li> <li>1 pc Arsenic, (As)</li> <li>1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.</li> <li>1 Set Air Filter</li> <li>1 Unit Silent Type Air Compressor</li> </ul>			
1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		Hallow Cathode Lamp	
1 pc Iron (Fe) 1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 pc Lead (Pb)	
1 pc Magnesium, (Mg) 1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		• ' '	
1 pc Manganese, (Mn) 1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor	L		
1 pc Cadmium, (Cd) 1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		<u> </u>	
1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 pc Manganese, (Mn)	
1 pc Chromium, (Cr) 1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 pc Cadmium, (Cd)	
1 pc Sodium, (Na) 1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
1 pc Zinc, (Zn) 1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
1 pc Arsenic, (As) 1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor	[	•	
1 bottle/each Standards for: Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 pc Arsenic, (As)	
Lead, Iron, Magnesium, Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor		1 bottle/each Standards for:	
Manganese, Cadmium, Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
Chromium, Sodium, Zinc.  1 Set Air Filter  1 Unit Silent Type Air Compressor			
1 Set Air Filter  1 Unit Silent Type Air Compressor			
1 Unit Silent Type Air Compressor		Chromium, Sodium, Zinc.	
1 Unit Silent Type Air Compressor			
Compressor	1	1 Set Air Filter	
Compressor			
Compressor		1 Unit Silent Type Air	
		* -	
1 Set Hood and Ducting works		Compressor	
1 Set Hood and Ducting works		10.77	
		1 Set Hood and Ducting works	
1 Unit Acetylene Gas		1 Unit Acetylene Gas	
		<b>,</b>	
1 Unit Acetylene Regulator"		1 Unit Acetylene Regulator"	
1 Onit rectytone regulator		1 Chit / teoryiche regulator	

After Sales Services: Guaranteed availability of supplies and spare parts for at least ten (10) years	
Free on-site training included	
Warranty: Two (2) years for parts and services; One (1) Preventive Maintenance within the warranty period	

# Section VIII. Checklist of Technical and Financial Documents

## **Notes on the Checklist of Technical and Financial Documents**

The prescribed documents in the checklist are mandatory to be submitted in the Bid, but shall be subject to the following:

- a. GPPB Resolution No. 09-2020 on the efficient procurement measures during a State of Calamity or other similar issuances that shall allow the use of alternate documents in lieu of the mandated requirements; or
- b. Any subsequent GPPB issuances adjusting the documentary requirements after the effectivity of the adoption of the PBDs.

The BAC shall be checking the submitted documents of each Bidder against this checklist to ascertain if they are all present, using a non-discretionary "pass/fail" criterion pursuant to Section 30 of the 2016 revised IRR of RA No. 9184.

# **Checklist of Technical and Financial Documents**

## I. TECHNICAL COMPONENT ENVELOPE

#### Class "A" Documents

Les	gal Do	cuments
	(a)	Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages);
	(b)	or Registration certificate from Securities and Exchange Commission (SEC), Department of Trade and Industry (DTI) for sole proprietorship, or Cooperative Development Authority (CDA) for cooperatives or its equivalent document,
	(c)	and Mayor's or Business permit issued by the city or municipality where the principal place of business of the prospective bidder is located, or the equivalent document for Exclusive Economic Zones or Areas; and
	(d)	Tax clearance per E.O. No. 398, s. 2005, as finally reviewed and approved by the Bureau of Internal Revenue (BIR).
Teo	chnica	l Documents
	(f)	Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid; <b>and</b>
	(g)	Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents; and
	(h)	Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;  or
	(i)	Original copy of Notarized Bid Securing Declaration; <u>and</u> Conformity with the Technical Specifications, which may include production/delivery schedule, manpower requirements, and/or aftersales/parts, if applicable; <u>and</u>
	(j)	Original duly signed Omnibus Sworn Statement (OSS); and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.
Fin	ancia	l Documents
	(k)	The Supplier's audited financial statements, showing, among others, the Supplier's total and current assets and liabilities, stamped "received" by the BIR or its duly accredited and authorized institutions, for the preceding calendar year which should not be earlier than two (2) years from the date of bid submission; <b>and</b>
	(l)	The prospective bidder's computation of Net Financial Contracting Capacity (NFCC);

		or A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation.
		Class "B" Documents
	(m)	If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence;
		or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.
Oth	ner do	cumentary requirements under RA No. 9184 (as applicable)
		[For foreign bidders claiming by reason of their country's extension of
	(0)	reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product. Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity.
FIN	NAN(	CIAL COMPONENT ENVELOPE
	(a)	Original of duly signed and accomplished Financial Bid Form; and
	(b)	Original of duly signed and accomplished Price Schedule(s).
		<u>Other do</u> ☐ (n) ☐ (o) <b>FINANC</b> ☐ (a)

# Annexes

Annex "A"

Statement of All Ongoing Government and Private Contracts

a. Date Awarded	b. Contract Effectivity	c. Date Completed					
a. Amount of Award	ınt of tion	c. Contract Duration					
		%					
Didda", Didd	Diddel 8 role	Description					
	Nature of Work						
a. Owner's Name	b. Address	c. Telephone Nos.					
	Name of Contract		Government:		Private:		

## Annex "B"

## **Statement of Single Largest Completed Contract**

a. Date Awarded	b. Contract Effectivity	c. Date Completed					
a. Amount of Award	b. Amount of Completion	c. Contract Duration					
		%					
Didda" Dola	בומוכן א ועסוכ	Description					
	Nature of Work						
a. Owner's Name	b. Address	c. Telephone Nos.					
	Name of Contract		Government:		Private:		

## **Bid Securing Declaration**

REPUBLIC OF THE PI	HILIPPINES)
CITY OF	) S.S.
X	

#### **BID SECURING DECLARATION**

**Invitation to Bid:** [Insert Reference number]

To: [Insert name and address of the Procuring Entity]

I/We<sup>3</sup>, the undersigned, declare that:

1. I/We understand that, according to your conditions, bids must be supported by a Bid Security, which may be in the form of a Bid-Securing Declaration.

- 2. I/We accept that: (a) I/we will be automatically disqualified from bidding for any contract with any procuring entity for a period of two (2) years upon receipt of your Blacklisting order; and, (b) I/we will pay the applicable fine provided under Section 6 of the Guidelines on the Use of Bid Securing Declaration, within fifteen (15) days from receipt of the written demand by the procuring entity for the commission of acts resulting to the enforcement of the bid securing declaration under Sections 23.1(b), 34.2, 40.1 and 69.1, except 69.1(f), of the IRR of RA 9184; without prejudice to other legal action the government may undertake.
- 3. I/We understand that this Bid Securing Declaration shall cease to be valid on the following circumstances:
  - (a) Upon expiration of the bid validity period, or any extension thereof pursuant to your request;
  - (b) I am/we are declared ineligible or post-disqualified upon receipt of your notice to such effect, and (i) I/we failed to timely file a request for reconsideration or (ii) I/we filed a waiver to avail of said right;
  - (c) I am/we are declared the bidder with the Lowest Calculated Responsive Bid, and I/we have furnished the performance security and signed the Contract.

 $<sup>^3</sup>$  Select one and delete the other. Adopt the same instruction for similar terms throughout the document.

IN WITNESS WHEREOF, I/We have hereunto set my/our hand/s this \_\_\_\_ day of [month] [year] at [place of execution].

[Insert NAME OF BIDDER'S AUTHORIZED REPRESENTATIVE]
[Insert Signatory's Legal Capacity]
Affiant

of execution], Philippines. Affiant/s is/are per me through competent evidence of identity as (A.M. No. 02-8-13-SC). Affiant/s exhibite identification card used], with his/her photog and his/her Community Tax Certifity  Witness my hand and seal this day	fore me this day of [month] [year] at [place sonally known to me and was/were identified by s defined in the 2004 Rules on Notarial Practice ed to me his/her [insert type of government graph and signature appearing thereon, with no. icate No issued on at  y of [month] [year].  NAME OF NOTARY PUBLIC Serial No. of Commission Notary Public for until Roll of Attorneys No PTR No [date issued], [place issued]  [BP No [date issued], [place issued]
Doc. No Page No Book No Series of	

#### **Omnibus Sworn Statement**

REPUBLIC OF THE PHILIPPINES	)	
CITY/MUNICIPALITY OF	);	S.S

#### **AFFIDAVIT**

I, [Name of Affiant], of legal age, [Civil Status], [Nationality], and residing at [Address of Affiant], after having been duly sworn in accordance with law, do hereby depose and state that:

#### 1. Select one, delete the other:

If a sole proprietorship: I am the sole proprietor or authorized representative of [Name of Bidder] with office address at [address of Bidder];

If a partnership, corporation, cooperative, or joint venture: I am the duly authorized and designated representative of [Name of Bidder] with office address at [address of Bidder];

#### 2. Select one, delete the other:

If a sole proprietorship: As the owner and sole proprietor, or authorized representative of [Name of Bidder], I have full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached duly notarized Special Power of Attorney;

If a partnership, corporation, cooperative, or joint venture: I am granted full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached [state title of attached document showing proof of authorization (e.g., duly notarized Secretary's Certificate, Board/Partnership Resolution, or Special Power of Attorney, whichever is applicable;)];

- 3. *[Name of Bidder]* is not "blacklisted" or barred from bidding by the Government of the Philippines or any of its agencies, offices, corporations, or Local Government Units, foreign government/foreign or international financing institution whose blacklisting rules have been recognized by the Government Procurement Policy Board;
- 4. Each of the documents submitted in satisfaction of the bidding requirements is an authentic copy of the original, complete, and all statements and information provided therein are true and correct:
- 5. [Name of Bidder] is authorizing the Head of the Procuring Entity or its duly authorized representative(s) to verify all the documents submitted;

6	Select	one	delete	the	rest
o.	$\mathcal{L}_{\mathcal{L}}}}}}}}}}$	oii	ucicic	u	1000.

If a sole proprietorship: The owner or sole proprietor is not related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

If a partnership or cooperative: None of the officers and members of [Name of Bidder] is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

If a corporation or joint venture: None of the officers, directors, and controlling stockholders of [Name of Bidder] is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

- 7. [Name of Bidder] complies with existing labor laws and standards; and
- 8. *[Name of Bidder]* is aware of and has undertaken the following responsibilities as a Bidder:
  - a) Carefully examine all of the Bidding Documents;
  - b) Acknowledge all conditions, local or otherwise, affecting the implementation of the Contract;
  - c) Made an estimate of the facilities available and needed for the contract to be bid, if any; and
  - d) Inquire or secure Supplemental/Bid Bulletin(s) issued for the [Name of the Project].
- 9. [Name of Bidder] did not give or pay directly or indirectly, any commission, amount, fee, or any form of consideration, pecuniary or otherwise, to any person or official, personnel or representative of the government in relation to any procurement project or activity.

IN WITNESS WHEREOF, I	have hereunto set my hand this day of, 20 at
, Philippines.	
11	
	Ridder's Representative/Authorized Signatory

**SUBSCRIBED AND SWORN** to before me this \_\_\_\_ day of [month] [year] at [place of execution], Philippines. Affiant/s is/are personally known to me and was/were identified by

(A.M. No. 02-8-13-SC). Affiant/s exhibited identification card used], with his/her photos	as defined in the 2004 Rules on Notarial Practice ted to me his/her [insert type of government ograph and signature appearing thereon, with no. ficate No issued on at
Witness my hand and seal this d	ay of [month] [year].
	NAME OF NOTARY PUBLIC  Serial No. of Commission  Notary Public for until  Roll of Attorneys No  PTR No [date issued], [place issued]  IBP No [date issued], [place issued]
Doc. No Page No Book No Series of	

<sup>\*</sup> This form will not apply for WB funded projects.

## **Bid Form**

	Invitation to	Date: o Bid <sup>4</sup> N <sup>o</sup> :	<u> </u>
To: [name and address of Proc	uring Entity]		
Gentlemen and/or Ladies:			
numbers], the receipt of which [supply/deliver/perform] [desc] Documents for the sum of [total	n is hereby duly activition of the Go all Bid amount in we	s including Bid Bulletin Numbers eknowledged, we, the undersigned, ods] in conformity with the said I ords and figures] or such other sums f Prices attached herewith and made	offer to Bidding as may
We undertake, if our Bid i schedule specified in the Sched	<u> </u>	ver the goods in accordance with the d	lelivery
If our Bid is accepted, v amounts, and within the times s	_	covide a performance security in the dding Documents.	e form,
•	source not found	alidity Period specified in <u>BDS</u> provises and it shall remain binding upon of that period.	
Commissions or gratuities and to contract execution if we		be paid by us to agents relating to to ontract, are listed below:5	his Bid,
Name and address of agent	Amount and Currency	Purpose of Commission or gratuity	
(if none, state "None")	-		

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof and your Notice of Award, shall be binding upon us.

We understand that you are not bound to accept the Lowest Calculated Bid or any Bid you may receive.

<sup>&</sup>lt;sup>4</sup> If ADB, JICA and WB funded projects, use IFB.

<sup>&</sup>lt;sup>5</sup> Applicable only if the Funding Source is the ADB, JICA or WB.

We certify/confirm that we comply with the eligibility requirements as per **ITB** Clause **Error! Reference source not found.** of the Bidding Documents.

We likewise certify/confirm that the undersigned, [for sole proprietorships, insert: as the owner and sole proprietor or authorized representative of <u>Name of Bidder</u>, has the full power and authority to participate, submit the bid, and to sign and execute the ensuing contract, on the latter's behalf for the <u>Name of Project</u> of the <u>Name of the Procuring Entity</u>] [for partnerships, corporations, cooperatives, or joint ventures, insert: is granted full power and authority by the <u>Name of Bidder</u>, to participate, submit the bid, and to sign and execute the ensuing contract on the latter's behalf for <u>Name of Project</u> of the <u>Name of the Procuring Entity</u>].

We acknowledge that failure to sign each and every page of this Bid Form, including the attached Schedule of Prices, shall be a ground for the rejection of our bid.							
Dated this	day of		_ 20				
[signature]		[in the capa	city of]				

Duly authorized to sign Bid for and on behalf of

### Price Schedule

N	ame of Bio	lder		I	Invitation to Bid <sup>6</sup> Number Page of					
1	2	3	4	5	6	7	8	9		
em	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDF (col 4 x 8)		

<sup>&</sup>lt;sup>6</sup> If ADB, JICA and WB funded projects, use IFB.

## For Goods Offered from Within the Philippines

	Name of Bidder					Invitation to Bid <sup>7</sup> Number Page of				
1	2	3	4	5	6	7	8	9	10	
Item	Description	Country of origin	Quantity	Unit price EXW per item	Transportation and Insurance and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)	
		·]			[in the	e capacity (	<u>ef]</u>		-	

Duly authorized to sign Bid for and on behalf of \_\_\_\_\_

 $<sup>^{7}</sup>$  If ADB, JICA and WB funded projects, use IFB.

